

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Emerson Process Management
Remote Automation Solutions
Unit 2, Park Farm Courtyard
Easthorpe, Malton YO17 6QX
United Kingdom

Measuring Instrument An **electronic gas volume conversion device (EVCD)** intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : FloBoss S600 or S600+
Manufacturer's mark or name : Emerson Process Management
Remote Automation Solutions
Conversion principle : PTZ
Ambient temperature range : -10 °C / +55 °C
Designed for : non-condensing humidity
Environmental classes : M1 / E2

The intended location for the instrument is "closed".

Further properties are described in the annexes:

- Description number T10152 revision 20;
- Documentation folder number T10152-8.

Valid until 20 February 2029

Initially issued 20 February 2009

Remark This revision replaces the previous revisions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
29 April 2024

Certification Board