

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	EUROMET ÖLÇÜ ALETLERİ LABORATUVAR VE TEST EKİPMANLARI DANIŞMANLIK İMALAT SANAYİ VE TİCARET LTD. ŞTİ. Anadolu Bulvarı ATB İş Merkezi C Blok No:84 Yenimahalle - Ankara Turkey
Measuring instrument	A water meter Trade mark : AYSU Type : GÜNHAN xx-yyy; AYHAN xx-yyy; YILDIZHAN xx-yyy; GÖKHAN xx-yyy; DAĞHAN xx-yyy; DENİZHAN xx-yyy; OĞUZHAN xx-yyy; AVŞAR xx-yyy; BAYINDIR xx-yyy Measuring Range : See paragraph 1.2 of the description. Environment classes : M1 / E1 Temperature range water : +0,1 °C / +50 °C (T50 class meters) +0,1 °C / +90 °C (T90 class meters) Temperature range ambient : -10 °C / +55 °C Destined for the measurement of : Water Further properties are described in the annexes: – Description T10667 revision 0; – Documentation folder T10667-1.
Valid until	2 October 2024

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
2 October 2014


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.