

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Zhejiang Chint Instrument & Meter Co., Ltd
Wenzhou Bridge
Industrial Zone Yueqing
CN 325603 Wenzhou Zhejiang
P.R. China

Measuring instrument **Diaphragm gas meter**
Type : ZT-G(value)S
G(value)S
G(value)A
(value = 1.6, 2.5 or 4)

Manufacturer's mark or name : Chint
Destined for the measurement of : Gas volume
Accuracy class : Class 1,5
Environment classes : M1 / E2
Temperature range : ZT-G(value)S : -10 °C / +40 °C
G(value)S : -25 °C / +55 °C
G(value)A : -25 °C / +55 °C
Location : Closed

Further properties are described in the annexes
– Description T10052 revision 6
– Documentation folder T10052-7

Valid until 13 November 2027

Remarks This revision replaces the earlier version(s), including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
27 February 2023

Certification Board