

Issued by	NMI Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:		
Manufacturer	Talamona S.r.l. Via Roma 772 23018 Talamona (SO) Italy		
Measuring instrument	A Turbine Gas Meter		
	Type	:	iMTM-CT
	Manufacturer's mark or name	:	Talamona S.r.l.
	Destined for the measurement of	:	Gas volume
	Accuracy class	:	Class 1,0
	Environment classes	:	M1 / E2
	Temperature range	:	-25 °C / +55 °C
	Further properties are described in the annexes: – Description T10080 revision 9; – Documentation folder T10080-3.		
Valid until	12 July 2027		
Remark	This revision replaces the earlier versions, including its documentation folder.		

Issuing Authority

NMI Certin B.V., Notified Body number 0122
9 October 2018



C. Oosterman
Head Certification Board

NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin BV.as Notified Body can be verified at
<http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.