

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:

Manufacturer Itron GmbH
Hardeckstrasse 2
D-76185 Karlsruhe
Germany

Measuring instrument **A Diaphragm Gas Meter**

Manufacturer : Itron GmbH or Actaris
Type : G65 and G100
Destined for the measurement of : Gas volume
Accuracy class : Class 1,5
Environment classes : M1 / E2
Temperature range : - 10 °C / +40 °C

G-value	Q _{max} [m ³ /h]	Q _t [m ³ /h]	Q _{min} [m ³ /h]	V [dm ³]
65	100	10	0,65	55
100	160	16	1	123

Further properties are described in:

- Description T10100 revision 1
- Documentation folder T10100-1

Valid until 7 January 2018

Remarks This revision was issued due to the name change of the manufacturer from Actaris to Itron GmbH and the addition of an optional Cyble Adapter. This revision replaces the earlier version, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
14 December 2009

C. Oosterman
Head of the Certification Board