

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Ex~i Flow Measurement Ltd.
Unit 22, Ford Lane Business Park, Ford Lane
FORD, West Sussex BN18 0UZ
United Kingdom

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : SFC3000
Manufacturer's mark or name : Ex~i Flow Measurement
Conversion principle : PTZ
Ambient temperature range : -25 °C / +55 °C
Designed for : non-condensing humidity
Environment classes : M2 / E2
The intended location for the instrument is "closed".

Further properties are described in the annexes:

- Description T10159 revision 5;
- Documentation folder T10159-5.

Valid until 7 January 2029

Initially issued 7 January 2009

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
31 August 2023

Certification Board