

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The "Metrologiewet" (Stb. 2006, 137) as Dutch implementation of Directive 2004/22/EC on measuring instruments (MID).

Manufacturer SAMGAS ROMANIA SRL
Strada Mierlei 12
Arad
Romania

In respect of **A diaphragm gas meter**

Manufacturer : SAMGAS Romania
Type : RS/2001/RO LA (steel housing)
RS/2001/RO AL (aluminium housing)

Characteristics Destined for the measurement of : gas volume
Accuracy class : Class 1,5
Temperature range : -10°C / +40°C
Mechanical environment class : M2
Electromagnetic environment class : E1

G-value	Q _{max} [m ³ /h]	Q _t [m ³ /h]	Q _{min} [m ³ /h]
4	6	0,60	0,040
2,5	4	0,40	0,025
1,6	2,5	0,25	0,016

In the description number T10253 Revision 0 further characteristics are described.

Valid until 26 November 2019

Description and Documentation The gas meter is described in the description number T10253 Revision 0 and documented in the Documentation folder T10253-1, appertaining to this EC-type examination certificate.

Notified Body number 0122
NMI Certin, 26 November 2009

C. Oosterman
Head Certification Board