

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Elster GmbH
Steinernstraße 19
55252 Mainz-Kastel
Germany

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be
used for gas volume conversion as a sub-assembly (according to article 4 of
the MID) of a gas meter.

Type : EK280

Manufacturer's mark or name : Elster

Conversion principle : T, PT or PTZ

Ambient temperature range : -25 °C / +55 °C

Designed for : condensing humidity

Environment classes : M2 / E2

The intended location for the instrument is open.

- Further properties are described in the annexes:
- Description T10339 revision 26;
 - Documentation folder T10339-11.

Initially issued 31 May 2011

Valid until 31 May 2031

Remark This version replaces the earlier versions; excluding its documentation
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
30 November 2023

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMI Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

This document is digitally signed and
sealed. The digital signature can be
verified in the blue ribbon at the top
of the electronic version of this
certificate.

Reproduction of the complete
document only is permitted.