

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 9 of Directive 2004/22/EC, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2004/22/EC, to:

Manufacturer Hangzhou Sintta meter technology Co.,Ltd.
Add:7# building
No. 31 Xianxing Road
Xianlin Town, Yuhang district 311122
Hangzhou, P.R. China

Measuring instrument **A Diaphragm Gas Meter**

Type : G1,6/2,5/4 S (Steel casing)
G1,6/2,5/4A (Aluminum casing)

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M2 / E1

Temperature range : -10°C / +40°C


Location : Closed

G-value	Q _{max} [m ³ /h]	Q _{min} [m ³ /h]	Q _t [m ³ /h]	V [dm ³]
4	6	0,040	0,6	1,2
2,5	4	0,025	0,4	1,2
1,6	2,5	0,016	0,25	1,2

Further properties are described in the annexes:
– Description T10445 revision 0;
– Documentation folder T10445-1.

Valid until 11 October 2023

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
11 October 2013



C. Oosterman
Head Certification Board