

Issued by	NM ⁱ Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Dandong Gas Meter Co.,Ltd. No. 66 Gucheng Road, Yuanbao District, Dandong City, Liaoning Province, People's Republic of China
Measuring instrument	A Diaphragm Gas Meter Type : J1.6CII : J2.5CII Destined for the measurement of : Gas volume Accuracy class : Class 1,5 Environment classes : M2 / E1 Temperature range : - 25 °C / +55 °C Location : Closed
	Further properties are described in the annexes – Description T10451 revision 2 – Documentation folder T10451-3
Valid until	19 June 2024
Remarks	This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMⁱ Certin B.V., Notified Body number 0122**
 15 August 2016



C. Oosterman
 Head Certification Board

NMⁱ Certin B.V.
 Hugo de Grootplein 1
 3314 EG Dordrecht
 The Netherlands
 T +31 78 6332332
 certin@nmi.nl
 www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMⁱ Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMⁱ (see www.nmi.nl).

Reproduction of the complete document only is permitted.