

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	CPL Concordia Soc. Coop. Via A. Grandi, 39 41033 Concordia (MO) Italy
Measuring instrument	An <b>electronic gas-volume conversion device (EVCD)</b> , intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter. Type : ECOR4 Conversion principle : T, PT or PTZ Device type : 1 (complete system) Ambient temperature range : -25 °C / +55 °C Designed for : condensing humidity Environment classes : M2 / E2 The intended location for the instrument is open.  Further properties are described in the annexes – Description T10452 revision 2 – Documentation folder T10452-3
Valid until	24 April 2022
Remarks	This revision replaces the earlier version, including its Documentation folder.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**  
28 February 2013

C. Oosterman  
Head Certification Board

**NMi Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see [www.nmi.nl](http://www.nmi.nl)).

Reproduction of the complete document only is permitted.