	E U - t y	уре	examination certificate
			Number <b>T10466</b> revision 10 Project number 3664283 Page 1 of 1
Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:		
Manufacturer	Flow Meter Group B.V. Innovatieweg 32 7007 CD Doetinchem The Netherlands		
Measuring instrument	A <b>Turbine Gas Meter</b> Type	-	FMT-L FMT-S FMT-Lx FMT-Dc
	Destined for the measurement Accuracy class Environment classes Temperature range	:	Gas volume Class 1,0 M1 / E1 -40 °C / +70 °C
(-	Further properties are described in the annexes – Description T10466 revision 10; – Documentation folder T10466-7.		
Valid until	5 July 2033		
Initially issued	5 July 2013		
Remarks	This revision replaces the earli folder.	er versions	, except for its documentation

Issuing Authority



NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 636 2332 certin@nmi.nl www.nmi.nl **NMi Certin B.V., Notified Body number 0122** 5 July 2023

## **Certification Board**

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <u>http://</u> ec.europa.eu/growth/tools-databases/nando/ Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.

