

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer S.C. AEM S.A.  
Calea Buziasului nr. 26  
Timisoara 300693  
Romania

Measuring instrument **Ultrasonic gas meter**

Type	: G4 EUS G4 EUS-2
Manufacturer's mark or name	: S.C. AEM S.A.
Destined for the measurement of	: Gas volume of natural gas, See description §1.2.1
Accuracy class	: Class 1,5
Environment classes	: M1 / E1
Temperature range	: - 25 °C / +55 °C
Designed for	: Condensing humidity
Location	: Open

Further properties are described in the annexes;  
– Description T10478 revision 16;  
– Documentation folder T10478-13.

Valid until 27 November 2032

Remark This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
17 February 2023

## Certification Board