

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Tancy Instrument Group Co., Ltd.
No. 198, Hualian Road, Cangnan Industrial Zone
Wenzhou City, Zhejiang Province
P.R. China

Measuring instrument **A Rotary Displacement Gas Meter**

Type	: TYL
Manufacturer's mark or name	: Tancy
Destined for the measurement of	: Gas volume
Accuracy class	: Class 1,0
Environment classes	: M2 / E1
Gas and ambient temperature range	: -25 °C / +55 °C

Further properties are described in the annexes:

- Description T10487 revision 7;
- Documentation folder T10487-4.

Initially issued 2 November 2012

Valid until 2 November 2032

Remark This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
20 April 2023

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.