





Number T10488 revision 3 Project number 2514194 Page 1 of 1

Issued by NMi Certin B.V.,

> designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable

requirements of Directive 2014/32/EU, to:

Manufacturer Tancy Instrument Group Co., Ltd.

No. 198, Hualian Road, Cangnan Industrial Zone

Wenzhou City, Zhejiang Province

P.R. China

Measuring instrument A Turbine Gas Meter

TBQM Type Manufacturer's mark or name Tancy

Destined for the measurement of Gas volume Accuracy class Class 1.0 **Environment classes** : M1/E1

-10 °C / +55 °C Gas temperature range Ambient temperature range -10 °C / +55 °C

Further properties are described in the annexes:

Description T10488 revision 3;

- Documentation folder T10488-2.

Valid until 26 September 2024

This revision replaces the earlier versions, except for its documentation



Issuing Authority

Remark





Certification Board

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at http:// ec.europa.eu/growth/tools-databases/nando/ Reproduction of the complete

sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.





NMi Certin B.V. Thijsseweg 11 2629 JA Delft The Netherlands T +31 88 636 2332 certin@nmi.nl www.nmi.nl

