

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer SICK Engineering GmbH  
Bergener Ring 27  
D-01458 Ottendorf-Okrilla  
Germany

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type : SICK Flow-X/M, SICK Flow-X/P,  
SICK Flow-X/S, SICK Flow-X/R and  
SICK Flow-X/K.

Manufacturer's mark or name : SICK

Conversion principle : PTZ

Ambient temperature range : -25 °C / +55 °C

Designed for : non-condensing humidity

Environment classes : M2 / E2

The intended location for the instrument is open.

Further properties are described in the annexes:

- Description T10548 revision 13;
- Documentation folder T10548-9.

Valid until 28 February 2033

Initially issued 28 February 2013

Remark

- This revision replaces the previous revisions.
- The documentation folder is not changed.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
8 December 2023

Certification Board