

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 9 of Directive 2004/22/EC, after  
having established that the Measuring instrument meets the applicable  
requirements of Directive 2004/22/EC, to:

Manufacturer CPL Concordia Soc. Coop.  
Via A. Grandi, 39  
41033 Concordia (MO)  
Italy

Measuring instrument **A Diaphragm Gas Meter**

Type : EG16 or EG16 SMART

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M2 / E1

Temperature range : - 25 °C / +55 °C


Location : Closed

G-value	Q <sub>max</sub> [m <sup>3</sup> /h]	Q <sub>min</sub> [m <sup>3</sup> /h]	Q <sub>t</sub> [m <sup>3</sup> /h]	V [dm <sup>3</sup> ]	p <sub>max</sub> [bar]
16	25	0,16	2,5	6	0,5

Further properties are described in the annexes:  
– Description T10588 revision 0;  
– Documentation folder T10588-1.

Valid until 6 September 2023

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
6 September 2013



C. Oosterman  
Head Certification Board

**NMI Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that  
no liability is accepted and that the applicant  
shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified  
Body can be verified at [http://  
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge  
objection against this decision,  
within six weeks after the date of  
submission, to the general manager  
of NMI (see [www.nmi.nl](http://www.nmi.nl)).

Reproduction of the complete  
document only is permitted.