

EU-type examination certificate

Number **T10605** revision 4 Project number SO16201376 Page 1 of 1

Issued by + + + NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable

requirements of Directive 2014/32/EU, to:

Manufacturer EDMI Electronics Sdn Bhd

PL0226A, Jalan Cyber 1A, Senai Industrial Park Phase 3

81400 Senai Johor

Malaysia

Measuring instrument An Ultrasonic Gas Meter

Type : Helios

Destined for the measurement of : Gas volume of natural gas,

type H, L, E

Accuracy class : Class 1,5 Environment classes : M1 / E2

Gas temperature range : - 25 °C / +55 °C

Ambient temperature range : - 25 °C / +55 °C

Designed for : Condensing humidity

Intended location + + + + + + : Open

Further properties are described in the annexes:

Description T10605 revision 4;Documentation folder T10605-4

Valid until + + + 7 March 2024

Remark This revision replaces the earlier versions, including its documentation

folder.

Issuing Authority NMI Certin B.V., Notified Body number 0122

23 May 2016

C. Oosterman

Head Certification Board

NMi Certin B.V.

Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at http://ec.europa.eu/enterprise/newapproach/nando/

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.

