

EC type-examination Certificate

Number **T10623** revision 0
Project number 14200142
Page 1 of 1

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 9 of Directive 2004/22/EC, after having established that the Measuring instrument meets the applicable requirements of Directive 2004/22/EC, to:
Manufacturer	Petrogas Gas Systems bv Harderwijkweg 2 2803 PW Gouda The Netherlands
Measuring instrument	An interruptible measuring system on a skid Type : UAN metering skid; NH3 metering skid Destined for the measurement of : UAN and ammonia (NH3) Accuracy class : 0.5 for UAN and 1.0 for NH3 Environment classes : M2 / E2 Temperature range liquid : -10 °C / +65 °C Temperature range ambient : -10 °C / +55 °C $Q_{min} - Q_{max}$: 14 – 150 t/h Minimum measured quantity : See § 1.2 of the description Further properties are described in the annexes: – Description T10623 revision 0; – Documentation folder T10623-1.
Valid until	13 July 2025
Remark	This measuring system is approved for measuring mass only.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
13 July 2015


C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified
Body can be verified at [http://
ec.europa.eu/enterprise/newapproach/nando/](http://ec.europa.eu/enterprise/newapproach/nando/)

Parties concerned can lodge
objection against this decision,
within six weeks after the date of
submission, to the general manager
of NMi (see www.nmi.nl).

Reproduction of the complete
document only is permitted.