

Issued by NMI Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Itron-Ganz Meter Company Ltd.  
Táncsics Mihály Utca 11,  
H 2100 Gödöllő,  
Hungary

Measuring instrument **A Diaphragm Gas Meter**

Type : RF1 s

Manufacturer's mark or name : Itron

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M2 / E2

Temperature range : See paragraph 1.2.3 of the description

Location : Closed

G-value	Q <sub>max</sub> [m <sup>3</sup> /h]	Q <sub>min</sub> [m <sup>3</sup> /h]	Q <sub>t</sub> [m <sup>3</sup> /h]	V [dm <sup>3</sup> ]
6	10	0,060	1	6
4	6	0,040	0,6	4
2,5	4	0,025	0,4	2,5
1,6	2,5	0,016	0,25	1,6

Further properties are described in the annexes  
– Description T10654 revision 4  
– Documentation folder T10654-3

Valid until 8 August 2024

Remarks This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
21 July 2017



C. Oosterman  
Head Certification Board