

Issued by

NMi Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 17 of Directive 2014/32/EU, after  
having established that the Measuring instrument meets the applicable  
requirements of Directive 2014/32/EU, to:

Manufacturer

Metreg Technologies GmbH  
Neckaraue 9  
71686 Remseck  
Germany

Measuring instrument

**A Turbine Gas Meter**  
Type : MTM  
Manufacturer's mark or name : Metreg  
Destined for the measurement of : Gas volume  
Accuracy class : Class 1,0  
Environment classes : M1 / E1  
Gas temperature range : -10 °C / +55 °C  
Ambient temperature range : -10 °C / +55 °C

Further properties are described in the annexes:

- Description T10660 revision 2;
- Documentation folder T10660-2.

Valid until

1 October 2024

Initially issued

1 October 2014

Remark

This revision replaces the earlier versions, including its documentation folder.

Issuing Authority

**NMi Certin B.V.**  
16 June 2023

Certification Board

**NMi Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.