

EU-type examination certificate

Number **T10706** revision 2 Project number 15200369 Page 1 of 1

Issued by Manufacturer	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to: Goldcard Smart Group Co.,Ltd. No 158, Jingiao Street, Economic and Technology development Zone Hangzhou City P.R. China			
Measuring instrument	An Electronic Thermal Gas Mete	er		
	Type + + + + + + + +	+ +	JGMS-G / JGMS-R	
	Manufacturer's mark or name	+ +	Goldcard Smart Group Co.,Ltd.	
	Destined for the measurement of	÷ ; ; ;	Gas volume of natural gas, type H or L	
	Accuracy class	44	Class 1,5	
	Environment classes	4.4	M1 / E2	
	Gas temperature range Ambient temperature range	÷÷.	- 25 °C / +55 °C - 25 °C / +55 °C	
	Designed for	44	Condensing humidity	
	Intended location	÷.	Closed + + + + + + + + + +	
		+ +		
	Further properties are described in – Description T10706 revision 2; – Documentation folder T10706-3.	the a	annexes: + + + + + + + + + + + + + + + + + + +	
+ Valid until	9 September 2026			
Remark	This revision replaces the previous folder.	versio	ons, including its documentation	
Issuing Authority	NMi Certin B.V., Notified Body r	numb	per 0122	
$\begin{array}{c} \bullet & \bullet \\ \bullet & \bullet &$	20 November 2017 C. Oosterman Head Certification Board			
NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands T +31 78 6332332 certin@nmi.nl www.nmi.nl	This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability. The designation of NMi Certin B.V. as Notified Body can be verified at http:// ec.europa.eu/growth/tools-databases/nando/		duction of the complete hent only is permitted. INSPECTION RVA 122	