

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Sensus UK Systems Ltd  
No 3 Lindenwood, Chineham Business Park  
Crockford Lane, Basingstoke, RG24 8QY  
United Kingdom

Measuring instrument **A Diaphragm Gas Meter**

Type : Cubix U6

Manufacturer's mark or name : Sensus UK Systems Ltd

Destined for the measurement of : Gas volume

Accuracy class : Class 1,5

Environment classes : M1 / E1

Temperature range : -10 °C / +40 °C

Location : Closed

G-value	Q <sub>max</sub> [m <sup>3</sup> /h]	Q <sub>min</sub> [m <sup>3</sup> /h]	Q <sub>t</sub> [m <sup>3</sup> /h]	V [dm <sup>3</sup> ]
4	6	0,025	0,6	2

Further properties are described in the annexes:  
– Description T10726 revision 0;  
– Documentation folder T10726-1.

Valid until 3 May 2026

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
3 May 2016

  
C. Oosterman  
Head Certification Board