

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	Tormene AG 748 Lachen 9428 Walzenhausen Switzerland
Measuring instrument	A Rotary Displacement Gas Meter Type : TAR TAR-D Destined for the measurement of : Gas volume Accuracy class : Class 1,0 or 1,5 (see table under 1.2.1 of the description) Environment classes : M1 / E2 Temperature range : -40 °C / +70 °C Further properties are described in the annexes: – Description T10784 revision 0; – Documentation folder T10784-1.
Valid until	13 May 2026

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
13 May 2016



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV.as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).

Reproduction of the complete document only is permitted.