

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Secure Meters Ltd.
Pratapnagar Industrial Area
Udaipur 313003
India

Measuring instrument **Thermal-mass flow gas meter**

Type	: Liberty Gas 200
Manufacturer's mark or name	: Secure
Destined for the measurement of	: Gas volume of natural gas, See description §1.2.2
Accuracy class	: Class 1,5
Environment classes	: M1 / E2
Temperature range	: - 25 °C / +55 °C
Designed for	: Condensing humidity
Location	: Closed

Further properties are described in the annexes:
– Description T11223 revision 13;
– Documentation folder T11223-8.

Valid until 11 October 2028

Remarks This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
24 May 2022

Certification Board