

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:
Manufacturer	RMFTEK Electronic Products and Systems Inc. 1368 Cad. No 18/76, İvedik OSB 06378, Ankara Turkey
Measuring instrument	An <b>electronic gas-volume conversion device (EVCD)</b> , intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter. Type : MICRO-Z family, types ZMAX, Z1, Z3 and Z5 Manufacturer's mark or name : RFLO Conversion principle : PTZ Device type : 2 (separate components) Ambient temperature range : -25 °C / +70 °C Designed for : condensing humidity Environment classes : M2 / E2 The intended location for the instrument is closed.  Further properties are described in the annexes: – Description T11509 revision 0; – Documentation folder T11509-1.
Valid until	24 December 2028

Issuing Authority **NMi Certin B.V., Notified Body number 0122**  
24 December 2018



C. Oosterman  
Head Certification Board

**NMi Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.