

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 17 of Directive 2014/32/EU, after  
having established that the Measuring instrument meets the applicable  
requirements of Directive 2014/32/EU, to:

Manufacturer Hermann Pipersberg jr. GmbH  
Felder Hof 2  
42899 Remscheid  
Germany

Measuring instrument **A Diaphragm Gas Meter with temperature conversion**

Type	: UG T
Manufacturer's mark or name	: PIPERSBERG
Destined for the measurement of	: Gas volume
Accuracy class	: Class 1,5
Environment classes	: M1 / E1
Ambient temperature range	: - 25 °C / +55 °C
Gas temperature range	: - 10 °C / +40 °C
Location	: Closed

Further properties are described in the annexes  
– Description T11719 revision 0  
– Documentation folder T11719-1

Valid until 2 October 2029

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
2 October 2019

  
C. Oosterman  
Head Certification Board