





Number **T12077** revision 4 Project number 3524833 Page 1 of 1

Issued by NMi Certin B.V.,

designated and notified by the Netherlands to perform tasks with respect to conformity asssessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets

the applicable requirements of Directive 2014/32/EU, to:

Manufacturer EDMI Limited

47 Yishun Industrial Park A

768724 SINGAPORE

Authorized EDMI Electronics Sdn Bhd

representative PL0226A, Jalan Cyber 1A, Senai Industrial Park Phase 3

81400 Senai Johor

Malaysia

Measuring instrument An Ultrasonic gas meter

Type : GR-8BU

Manufacturer's mark or name : EDMI

Destined for the measurement of : Gas flow of natural gas. See

description §1.2.3

Accuracy class : Class 1,5 Environment classes : M1 / E2

Gas temperature range :  $-10 \,^{\circ}\text{C} / +55 \,^{\circ}\text{C}$ Ambient temperature range :  $-10 \,^{\circ}\text{C} / +55 \,^{\circ}\text{C}$ 

Designed for : Condensing humidity

Further properties are described in the annexes:

- Description T12077 revision 4;

- Documentation folder T12077-1.

Valid until 6 May 2031

Remark This revision replaces the previous versions, except for its documentation

folder.

Issuing Authority NMi Certin B.V., Notified Body number 0122

14 June 2022

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl

www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified

Body can be verified at http:// ec.europa.eu/growth/tools-databases/nando/ Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.





