



EU - type examination certificate

Number **T12605** revision 0
Project number 2477621
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Insight Metering Systems, LLC
427 Prospect Ave SE
Grand Rapids, MI 49503
Unites States of America

Measuring instrument

An **ultrasonic gas meter**
Type

: iSonic

Manufacturer's mark or name

: insight

Destined for the measurement of

: Gas volume

Accuracy class

: Class 1,0

Environment classes

: M1 / E2

Ambient temperature range

: -40 °C / +70 °C

Gas temperature range

: -40 °C / +70 °C

Designed for

: Condensing humidity

Intended location

: Open

Further properties are described in the annexes:

- Description T12605 revision 0:
- Documentation folder T12605-1.

Valid until

27 October 2033

Initially issued

27 October 2023

Issuing Authority

NMi Certin B.V., Notified Body number 0122
27 October 2023

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at [http://](http://ec.europa.eu/growth/tools-databases/nando/)
ec.europa.eu/growth/tools-databases/nando/

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.

