

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity modules mentioned in article 17 of Directive 2014/32/EU, after having established that the Measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Emerson Process Management
Remote Automation Solutions
Unit 2, Park Farm Courtyard
Easthorpe, Malton YO17 6QX
United Kingdom

Measuring
Instrument

An **electronic gas volume conversion device (EVCD)** intended to be used for gas volume conversion as a sub-assembly (according to article 4 of the MID) of a gas meter.

Type	: FloBoss S600 or S600+
Manufacturer's mark or name	: Emerson Process Management Remote Automation Solutions
Conversion principle	: PTZ
Ambient temperature range	: -10 °C / +55 °C
Designed for	: non-condensing humidity
Environmental classes	: M1 / E2

The intended location for the instrument is "closed".

Further properties are described in the annexes:

- Description number T10152 revision 21;
- Documentation folder number T10152-8.

Valid until

20 February 2029

Initially issued

20 February 2009

Remark

This revision replaces the previous revisions, except for its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
7 October 2025

Certification Board