

Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity modules mentioned in article 17 of Directive 2014/32/EU, after  
having established that the Measuring instrument meets the applicable  
requirements of Directive 2014/32/EU, to:

Manufacturer Emerson Process Management  
Remote Automation Solutions  
Unit 2, Park Farm Courtyard  
Easthorpe, Malton YO17 6QX  
United Kingdom

Measuring Instrument An **electronic gas volume conversion device (EVCD)** intended to be  
used for gas volume conversion as a sub-assembly (according to article 4 of  
the MID) of a gas meter.

Type : FloBoss S600 or S600+  
Manufacturer's mark or name : Emerson Process Management  
Remote Automation Solutions  
Conversion principle : PTZ  
Ambient temperature range : -10 °C / +55 °C  
Designed for : non-condensing humidity  
Environmental classes : M1 / E2

The intended location for the instrument is "closed".

Further properties are described in the annexes:

- Description number T10152 revision 22;
- Documentation folder number T10152-8.

Valid until 20 February 2029

Initially issued 20 February 2009

Remark This revision replaces the previous revisions, except for its documentation  
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
7 May 2026

Certification Board

**NMI Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that no  
liability is accepted and that the manufacturer shall  
indemnify third-party liability.

The designation of NMI Certin B.V. as Notified Body  
can be verified at [http://ec.europa.eu/growth/tools-  
databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete  
document only is permitted.

This document is digitally  
signed and sealed. The digital  
signature can be verified in the  
blue ribbon at the top of the  
electronic version of this  
certificate.