

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer KROHNE Oil & Gas
Minervum 7441
4817 ZG Breda
The Netherlands

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be
used for gas volume conversion as a sub-assembly (according to article 4 of
the MID) of a gas meter.

Type : Summit 8800
Manufacturer's mark or name : KROHNE Oil & Gas
Conversion principle : PTZ
Ambient temperature range : -25 °C / +55 °C
Designed for : non-condensing humidity
Environment classes : M2 / E2
The intended location for the instrument is "closed".

Further properties are described in the annexes:

- Description T10185 revision 25;
- Documentation folder T10185-7.

Valid until 3 January 2029

Initially issued 2 February 2009

Remark This revision replaces the earlier versions, except for its documentation
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
7 March 2024

Certification Board