



# EU-type examination certificate

Number **T10335** revision 20  
Project number 4031533  
Page 1 of 1

Issued by

NMi Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the Measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Elster GmbH  
Steinern Straße 19-21  
55252 Mainz-Kastel  
Germany

Measuring instrument

An **Ultrasonic Gas Meter** with optionally volume conversion

Type	:	Q.Sonic <sup>plus</sup>
Destined for the measurement of	:	Gas volume
Accuracy class	:	Class 1,0
Environment classes	:	M1 / E2
Gas temperature	:	-40 °C ... +100 °C -30 °C ... +80 °C with EVCD
Ambient temperature	:	-40 °C ... +70 °C -40 °C ... +55 °C with EVCD
Designed for	:	Condensing humidity
Intended location	:	Open

Further properties are described in the annexes:

- Description T10335 revision 20;
- Documentation folder T10335-11.

Valid until

29 November 2031

Initially issued

29 November 2011

Remark

This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority

**NMi Certin B.V., Notified Body number 0122**

15 December 2025

Certification Board

**NMi Certin B.V.**  
Thijsseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision  
that no liability is accepted and that the  
manufacturer shall indemnify third-party  
liability.

The designation of NMi Certin B.V. as  
Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete  
document only is permitted.

This document is digitally signed and  
sealed. The digital signature can be  
verified in the blue ribbon at the top  
of the electronic version of this  
certificate.

  
PRODUCTS  
RvA C 681