

EU-type examination certificate

Number **T10382** revision 9
Project number 4046641
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Apator Metrix S.A.
Grunwaldzka 14
83-110 Tczew
Poland

Measuring instrument

Diaphragm gas meter

Type	: UG T UG T HybridSmart
Destined for the measurement of	: Gas volume
Accuracy class	: Class 1,5
Environment classes	: M1 / E1
Temperature range	: -25 °C / +55 °C
Gas temperature range	: -25 °C / +40 °C
Location	: Closed

Further properties are described in the annexes:

- Description T10382 revision 9;
- Documentation folder T10382-7;
- Evaluation Report ER10382 revision 9.

Valid until

9 September 2031

Initially issued

9 September 2011

Remark

This revision replaces the earlier versions, except for its documentation
folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
5 October 2026

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at
<https://webgate.ec.europa.eu/single-market-compliance-space/home>

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.