



EU-type examination Certificate

Number **T10417** revision 10
Project number 4031571
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

RMG Messtechnik GmbH
Otto-Hahn-Strasse 5
D-35510, Butzbach
Germany

Measuring instrument

A Turbine Gas Meter
Type : TRZ03 and TRZ04
Manufacturer's mark or name : RMG Messtechnik
Destined for the measurement of : Gas volume
Accuracy class : Class 1,0
Environment classes : M1 / E2
Temperature range : -25 °C / +55 °C

Further properties are described in the annexes:

- Description T10417 revision 10;
- Documentation folder T10417-9.

Valid until

19 March 2032

Initially issued

19 March 2012

Remark

This revision replaces the earlier versions, including its documentation
folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122
4 November 2025

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The appointment of NMi Certin UK (TIC) Ltd.,
as Approved Body can be verified at
[https://www.gov.uk/uk-market-conformity-
assessment-bodies](https://www.gov.uk/uk-market-conformity-assessment-bodies)

Reproduction of the complete
document only is permitted.

This document is digitally signed and
sealed. The digital signature can be
verified in the blue ribbon at the top
of the electronic version of this
certificate.

