



EU-type examination Certificate

Number **T10434** revision 14
Project number 4018453
Page 1 of 1

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity modules mentioned in article 17 of Directive 2014/32/EU, after
having established that the Measuring instrument meets the applicable
requirements of Directive 2014/32/EU, to:

Manufacturer Elster GmbH
Steinern Straße 19-21
55252 Mainz-Kastel
Germany

Measuring instrument An **electronic gas-volume conversion device (EVCD)**, intended to be
used for gas volume conversion as a sub-assembly (according to article 4 of
the MID) of a gas meter.

Type : enCore FC1

Manufacturer's mark or name : Elster

Conversion principle : PT or PTZ

Ambient temperature range : -10 °C / +55 °C

Designed for : non-condensing humidity

Environment classes : M2 / E2

The intended location for the instrument is closed.

Further properties are described in the annexes:

- Description T10434 revision 14;
- Documentation folder T10434-12.

Valid until 25 June 2032

Initially issued 25 June 2012

Remark This revision replaces the previous versions, including its documentation
folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
26 September 2025

Certification Board

NMI Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMI Certin B.V. as
Notified Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed and
sealed. The digital signature can be
verified in the blue ribbon at the top
of the electronic version of this
certificate.

