



EU-type examination certificate

Number **T10586** revision 11
Project number 4031533
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Elster GmbH
Steinern Straße 19-21
55252 Mainz-Kastel
Germany

Measuring instrument

An **Ultrasonic Gas Meter** with optionally volume conversion

Type	:	Q.Sonic ^{max} ; Q.Sonic ^{max6}
Destined for the measurement of	:	Gas volume
Accuracy class	:	Class 1,0
Environment classes	:	M1 / E2
Gas temperature	:	-40 °C ... +100 °C -30 °C ... +80 °C with EVCD
Ambient temperature	:	-40 °C ... +70 °C -40 °C ... +55 °C with EVCD
Designed for	:	Condensing humidity
Intended location	:	Open

Further properties are described in the annexes:

- Description T10586 revision 11;
- Documentation folder T10586-5.

Valid until

6 May 2026

Initially issued

6 May 2016

Remark

This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122

15 December 2025

Certification Board

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

Reproduction of the complete
document only is permitted.

The designation of NMi Certin B.V. as
Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

This document is digitally signed and
sealed. The digital signature can be
verified in the blue ribbon at the top
of the electronic version of this
certificate.

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

RvA
PRODUCTS
RvA C 681