

Issued by

NMi Certin B.V.
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Metreg Technologies GmbH
Neckaraue 9
71686 Remseck
Germany

Measuring instrument

A Turbine Gas Meter
Type : MTM
Manufacturer's mark or name : Metreg
Destined for the measurement of : Gas volume
Accuracy class : Class 1,0
Environment classes : M1 / E1
Gas temperature range : -10 °C / +55 °C
Ambient temperature range : -10 °C / +55 °C

Further properties are described in the annexes:

- Description T10660 revision 4;
- Documentation folder T10660-2.

Valid until

1 October 2034

Initially issued

1 October 2014

Remark

This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority

NMi Certin B.V.
4 May 2026

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
<http://www.nmi.nl/>

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.