



EU-type examination certificate

Number **T10859** revision 2

Project number 4000957

Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Qianwei Meters (Chongqing) Co., Ltd.
Middle Section No.69, Huang Shan Avenue, Yu Bei District
401121 Chongqing
P.R. China

Measuring instrument

Diaphragm gas meter

Type

: G1.6
G2.5

Manufacturer's mark or name

: Qianwei

Destined for the measurement of

: Gas volume

Accuracy class

: Class 1,5

Environment classes

: M1 / E2

Temperature range

: - 10 °C / +40 °C

Designed for

: Condensing humidity

Location

: Open

Valid until

27 July 2027

Initially issued

27 July 2017

Remarks

This revision replaces the earlier version(s), including its documentation folder.

Issuing Authority

NMi Certin B.V., Notified Body number 0122

22 December 2025

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/hando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.


PRODUCTS
RvA C 681