

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Chengdu Qinchuan IoT Technology Co., Ltd.
No. 931, South 4th Road, Economic and Technological Development Zone,
Longquanyi District
610100 Chengdu
P.R. China

Measuring instrument **Diaphragm gas meter**
Type : G4-xx
G2.5-xx
G1.6-xx
Manufacturer's mark or name : Chengdu Qinchuan IoT Technology Co., Ltd.

Destined for the measurement of : Gas volume
Accuracy class : Class 1,5
Environment classes : M1 / E1
Temperature range : - 10 °C / +55 °C
Location : Closed

Further properties are described in the annexes.
– Description T10885 revision 4
– Documentation folder T10885-3

Valid until 18 January 2028

Initially issued 18 January 2018

Remarks This revision replaces the earlier versions, including for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
30 July 2024

Certification Board