

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Zenner Gas S.r.l.
Via Aterno, 122
66020 San Giovanni Teatino (CH)
Italy

Measuring instrument **Diaphragm gas meter**

Type	: BK04ZENNER01
Manufacturer's mark or name	: ZENNER
Destined for the measurement of	: Gas volume
Accuracy class	: Class 1,5
Environment classes	: M1 / E2
Temperature range	: - 25 °C / +55 °C
Designed for	: Condensing humidity
Location	: Closed

Further properties are described in the annexes:
– Description T11199 revision 13;
– Documentation folder T11199-4.

Valid until 7 August 2028

Initially issued 7 August 2018

Remark This revision replaces the earlier versions, except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
19 March 2024

Certification Board