

Issued by

NMi Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the Measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Endress+Hauser SICK GmbH+Co. KG  
Bergener Ring 27  
01458 Ottendorf-Okrilla  
Germany

Measuring instrument

An **electronic gas-volume conversion device (EVCD)**, intended to be  
used for gas volume conversion as a sub-assembly (according to article 4 of  
the MID) of a gas meter.

Type : Flow-X/C

Conversion principle : PTZ

Manufacturer's mark or name : Endress+Hauser

**Endress+Hauser** 

Ambient temperature range : -25 °C / +55 °C

Designed for : non-condensing humidity

Environment classes : M2 / E2

The intended location for the instrument is "closed".

Further properties are described in the annexes:

- Description T11449 revision 14;
- Documentation folder number T11449-8.

Valid until

24 December 2028

Initially issued

24 December 2018

Remark

This revision replaces the earlier versions, except for its documentation  
folder.

Issuing Authority

**NMi Certin B.V., Notified Body number 0122**  
30 September 2025

Certification Board

**NMi Certin B.V.**  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision  
that no liability is accepted and that the  
manufacturer shall indemnify third-party  
liability.

The appointment of NMi Certin UK (TIC) Ltd.,  
as Approved Body can be verified at  
[https://www.gov.uk/uk-market-conformity-](https://www.gov.uk/uk-market-conformity-assessment-bodies)  
assessment-bodies

Reproduction of the complete  
document only is permitted.

This document is digitally signed and  
sealed. The digital signature can be  
verified in the blue ribbon at the top  
of the electronic version of this  
certificate.