

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer Goldcard Smart Group Co., Ltd.
No. 158 Jinqiao Street
Qiantang District, Hangzhou
Zhejiang 310018
P.R. China

Measuring instrument **Diaphragm gas meter**
Type : JGD1.6S-M
JGD2.5S-M
JGD4S-M

Manufacturer's mark or name : Goldcard Smart Group Co., Ltd.

Destined for the measurement of : Gas volume
Accuracy class : Class 1,5
Environment classes : M1 / E2
Temperature range : -25 °C / +55 °C
Location : Closed
Type : JGD1.6S-M
JGD2.5S-M
JGD4S-M

Further properties are described in the annexes
– Description T11826 revision 4
– Documentation folder T11826-4

Valid until 19 June 2030

Initially issued 19 June 2020

Remarks This revision replaces the earlier versions, including its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
9 April 2024

Certification Board