


Issued by NMI Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer EDMi Limited  
47 Yishun Industrial Park A  
Singapore 768724

Authorized representative EDMi Electronics Sdn Bhd  
PL0226A, Jalan Cyber 1A, Senai Industrial Park Phase 3  
81400 Senai Johor  
Malaysia

Measuring instrument **Ultrasonic gas meter**

|                                 |  |
|---------------------------------|--|
| Type                            | : GC-25AU  |
| Manufacturer's mark or name     | :  EDMi |
| Destined for the measurement of | : Gas volume of natural gas, see description §1.2.2  |
| Accuracy class                  | : Class 1,5  |
| Environment classes             | : M1 / E2  |
| Temperature range               | : - 10 °C / +55 °C   |
| Designed for                    | : Condensing humidity  |
| Location                        | : Open   |

Further properties are described in the annexes:  
– Description T12198 revision 2;  
– Documentation folder T12198-2.

Valid until 7 June 2032

Initially issued 7 June 2022

Remark This revision replaces the earlier version(s), except for its documentation folder.

Issuing Authority **NMI Certin B.V., Notified Body number 0122**  
17 September 2024

Certification Board