

EU-type examination certificate

Number **T12651** revision 0
Project number 3651952
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

MeterSit
Viale dell'Industria, 31-33
35129 Padova
Italy

Measuring instrument

Thermal-mass flow gas meter

Type : NSM G4, NSM G6
Manufacturer's mark or name : MeterSit
Destined for the measurement of : Gas volume of natural gas,
See description §1.2.2
Accuracy class : Class 1,5
Environment classes : M1 / E2
Temperature range : - 25 °C / +55 °C
Designed for : Condensing humidity
Location : Open

Further properties are described in the annexes:

- Description T12651 revision 0;
- Documentation folder T12651-1.

Valid until

7 March 2034

Initially issued

7 March 2024

Issuing Authority

NMi Certin B.V., Notified Body number 0122
7 March 2024

Certification Board

NMi Certin B.V.
Thijssseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMi Certin B.V. as Notified
Body can be verified at [http://](http://ec.europa.eu/growth/tools-databases/nando/)
ec.europa.eu/growth/tools-databases/nando/

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.