

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Capital Innotech Private Ltd.
E-100, Mewar Industrial Area
Udaipur - 313003 Rajasthan
India

Measuring instrument

A Rotary Displacement Gas Meter

Type : CAP
Manufacturer's mark or name : Capital
Destined for the measurement of : Gas volume
Accuracy class : Class 1,0
Environment classes : M1 / E1
Temperature range : -25 °C / +55 °C

Further properties are described in the annexes:

- Description T12738 revision 0;
- Documentation folder T12738-1.

Valid until

17 May 2034

Initially issued

17 May 2024

Issuing Authority

NMi Certin B.V., Notified Body number 0122
17 May 2024

Certification Board