



EU-type examination certificate



Number **T12885** revision 0

Project number 3942966

Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in article 17 of Directive
2014/32/EU, after having established that the Measuring instrument meets
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Tancy Instrument Group Co., Ltd.
No. 198, Hualian d, Cangnan Industrial Park
Wenzhou City, Zhejiang Province
P.R. China

Measuring instrument

An **Ultrasonic Gas Meter**

Type

: TUF series

Manufacturer's mark or name

: Tancy

Destined for the measurement of

: Gas volume

Accuracy class

: Class 1,0

Environment classes

: M2 / E2

Ambient temperature range

: -25 °C / +55 °C

Designed for

: Condensing humidity

Intended location

: Open

Further properties are described in the annexes:

- Description T12885 revision 0;
- Documentation folder T12885-1.

Valid until

29 December 2035

Initially issued

29 December 2025

Issuing Authority

NMi Certin B.V., Notified Body number 0122

29 December 2025

Certification Board

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

Reproduction of the complete
document only is permitted.

The designation of NMi Certin B.V. as Notified
Body can be verified at <http://ec.europa.eu/growth/tools-databases/hando/>

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon
at the top of the electronic version
of this certificate.

