

EU-type examination certificate

Number **T12945** revision 0
Project number 3651952
Page 1 of 1

Issued by

NMi Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in article 17 of Directive 2014/32/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Italgas Reti S.p.A.
Largo Regio Parco 11
10153 Torino
Italy

Measuring instrument

Thermal-mass flow gas meter

Type

: NSM G4
NSM G6

Manufacturer's mark or name

: Italgas

Destined for the measurement of

: Gas volume of natural gas,
See description §1.2.2 Gas volume
of natural gas,
See description §1.2.2

Accuracy class

: Class 1,5

Environment classes

: M1 / E2

Temperature range

: - 25 °C / +55 °C

Designed for

: Condensing humidity

Location

: Open

Further properties are described in the annexes

- Description T12945 revision 0
- Documentation folder T12945-1

Valid until

13 December 2034

Initially issued

13 December 2024

Issuing Authority

NMi Certin B.V., Notified Body number 0122
13 December 2024

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 636 2332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.