

# EU-type examination certificate

Number **T13265** revision 0  
Project number 4128217  
Page 1 of 1

Issued by

NMi Certin B.V.,  
designated and notified by the Netherlands to perform tasks with respect to  
conformity assessment procedures mentioned in article 17 of Directive  
2014/32/EU, after having established that the measuring instrument meets  
the applicable requirements of Directive 2014/32/EU, to:

Manufacturer

Globaltronics For Electronics  
120 CPC Industrial park, 6 October City  
Giza  
Egypt

Measuring instrument

A **Diaphragm gas meter**

|                                 |                    |
|---------------------------------|--------------------|
| Type                            | : GT-SMART-PPG2.5  |
| Manufacturer's mark or name     | : Globaltronics    |
| Destined for the measurement of | : Gas volume       |
| Accuracy class                  | : Class 1,5        |
| Environment classes             | : M1 / E2          |
| Temperature range               | : - 25 °C / +55 °C |
| Location                        | : Closed           |

Further properties are described in the annexes:

- Description T13265 revision 0;
- Documentation folder T13265-1;
- Evaluation Report ER13265 revision 0.

Valid until

30 July 2036

Initially issued

30 July 2026

Issuing Authority

**NMi Certin B.V., Notified Body number 0122**  
30 July 2026

Certification Board

**NMi Certin B.V.**  
Thijssseweg 11  
2629 JA Delft  
The Netherlands  
T +31 88 636 2332  
[certin@nmi.nl](mailto:certin@nmi.nl)  
[www.nmi.nl](http://www.nmi.nl)

This document is issued under the provision that  
no liability is accepted and that the  
manufacturer shall indemnify third-party  
liability.

The designation of NMi Certin B.V. as Notified  
Body can be verified at  
<https://webgate.ec.europa.eu/single-market-compliance-space/home>

Reproduction of the complete  
document only is permitted.

This document is digitally signed  
and sealed. The digital signature  
can be verified in the blue ribbon at  
the top of the electronic version of  
this certificate.