

## EU-TYPE EXAMINATION CERTIFICATE

Certificate number	1073-17
Manufacturer	Shenzhen Star Instrument Co., Ltd.
Contact person	Mrs Li Wang
Address	Star Industry Park, Baolong Industry City, Longgang District
Postal code, City	518116, Shenzhen
Country	China
Production site	Shenzhen Star Instrument Co., Ltd.
Address	Star Industry Park, Baolong Industry City, Longgang District
Postal code, City	518116, Shenzhen
Country	China
Instrument	Electronic single-phase two-wire, two-element energy meter Direct connected
Mark - Type	STK106, DDSY23S, DDSY23III, DDS26D, STC106, DDSY23III, DDSYF23B, DDS26D, STZ106, DDSY23III, DDS26D
Register	LCD
Accuracy Class	Active: 1/C; Reactive: 2
Measurement range	100-127 V and 220-240 V, 0,15..5(100) A, 50 and 60 Hz Active: 1000 imp./kWh; Reactive: 1000 imp./kvarh
Temperature range	-40 .. 70 °C
Use	Outdoor
Protection Class	II
Environmental class	M1, E1 and E2
Registry method	Bidirectional method with always positive register: the meter always counts the energy of the measuring point as received energy, irrespective of the real energy direction

The energy meters meet the requirements of Directive 2014/32/EU, on the harmonisation of the laws of Member States relating to the making available on the market of measuring instruments (recast).

Certification was based on compliance with the following harmonised standards:

EN 50470-1 (2006)	Electricity metering equipment (a.c.)-part 1: General requirements, tests and test conditions - Metering equipment (class indexes A, B and C)
EN 50470-3 (2006)	Electricity metering equipment (a.c.)-part 3: Particular requirements - Static meters for active energy (class indexes A, B and C)
Valid until	13 July 2027

Based on a non-recurrent examination.

Version 1 of this certificate. The results are recorded in the following annex: test report 1289-17.

DNV GL Netherlands B.V.



J.P. Fonteijne  
Executive Vice President  
KEMA Laboratories



Arnhem, 13 July 2017

Copyright: Only integral reproduction of this Certificate is permitted without written permission from DNV GL.